




# FINEVISION

## Trifocal Hydrophilic

### Description



Model	POD F	
Material	26% Hydrophilic Acrylic	
Overall diameter	11.40mm	
Optic diameter	6.00mm	
Optic	Biconvex Aspheric Trifocal	
Haptic design	Double C-loop & Posterior Angulated Haptic	
Filtration	UV & Blue Light	
Refractive index	1.46	
Abbe number	58	
Additional power (IOL plane)	+1.75D & +3.50D	
Injection system	Medical Accuject 2.0 up to 24.5D Medical Accuject 2.1/2.2 up to 35D	
Spherical power	+6D to +35D (0.5D steps)	
Suggested A constant <sup>1</sup>	<b>Interferometry</b>	
	Hoffer Q: pACD	5.59
	Holladay 1: Sf	1.83
	Barrett: LF	1.86
	SRK/T: A	118.95
	Haigis <sup>2</sup> : a0; a1; a2	1.36; 0.4; 0.1

<sup>1</sup> Values estimated only; surgeons are recommended to personalize their A-constant based on their surgical techniques and equipment, experience with the lens model and postoperative results.

<sup>2</sup> Not optimized.

Note: The FINEVISION intraocular lens is not FDA approved.

**Contact Information:**  
[www.bvimedical.com/customer-support/](http://www.bvimedical.com/customer-support/)

[bvimedical.com](http://bvimedical.com)

## Product Information

<b>Manufacturer</b>	PhysIOL s.a. - Liège Science Park Allée des Noisetiers 4 B-4031 Belgium +32 4 361 05 49 physiol@bvimedical.com
<b>Certificate information</b>	CE (EU) 2017/745, Annex IX Chapter II : MDR 735733 R000 QMS (EU) 2017/745, Annex IX Chapter I and III : MDR 735719 R000 ISO 13485:2016 & EN ISO 13485:2016 : MD 658518 ISO 13485:2016 : MDSAP 691544
<b>Shelf life</b>	Five (5) years from manufacturing date
<b>Intended purpose</b>	Intended use (for all IOLs): The posterior chamber intraocular lens which is intended to be placed into the capsular bag for the replacement of the human lens to achieve the visual correction of aphakia in adult patients in whom the cataractous lens has been removed.
<b>Indication for use</b>	The lens should be used as intended in patients surgically treated for cataract, with possibly associated presbyopia, who desire improved uncorrected far vision, useful near and intermediate visual functions and reduced spectacle dependence.
<b>Product Composition</b>	No products of animal or human origin are present in the implant. The intraocular lens is 100% composed of the covalently crosslinked medical quality material HELIOFLEX, which is a (2-hydroxyethylmethacrylate; methylmethacrylate) copolymer including a UV and a blue light-filtering chromophores covalently bound to the material.
<b>For sterile product</b>	All IOLs from PhysIOL are steam sterilized
<b>Packaging Material</b>	Holder (Polypropylene) Container (Polypropylene) Storage liquid (0.9% NaCl solution) Aluminium lid (Aluminium Gold) Container label (paper) Blister PP (Polypropylene) Tyvek lid
<b>Product Class</b>	Classified as Class IIb implantable long-term surgically invasive medical devices under Rule 8 of Annex VIII of MDR 2017/745. Not available in the United States

**CE**  
2797

**Contact Information:**  
[www.bvimedical.com/customer-support/](http://www.bvimedical.com/customer-support/)



BVI and all other trademarks (unless noted otherwise) are property of BVI. BVI ©2022

590641-02